

ENABLING RELIABLE DIGITAL HOSPITALS

Michele Tomasin

Key Account Director, Healthcare & Pharmaceutical Segment Europe, Middle East & Africa

Contents

Executive Summary	3
Introduction	3
Diagnostic and Treatment Applications	
Modalities	4
In Vitro Diagnostic Laboratories	5
Field Technical Details	5
Hybrid Operating Theatres	6
Technical Recommendations	6
Picture Archiving and Communication System (PACS)	
Field Technical Details	8
Technical Recommendations	8
Radiology Information Systems (RIS) and Hospital Information Systems (HIS)	
Field Technical Details	9
Technical Recommendations	10
Conclusion	10



Executive Summary

Digital Imaging and Communications in Medicine (DICOM) is the standard for the transmission and storing of digital medical images. This method allows physicians to store, exchange and transmit medical images from devices such as scanners to printers and computers, allowing for the integration of medical imaging equipment with other devices. All relevant systems are increasingly becoming connected to Picture Archiving and Communications Systems (PACS), Radiology Information Systems (RIS), Hospital Information Systems (HIS), as well as to the hospital intranet and wider Internet. A physical infrastructure which is unable to support these applications can result in unexpected downtime and non-compliance. The following white paper will help to properly size the physical infrastructure when medical imaging and diagnostic equipment will be used, with a focus on power and cooling systems.

Introduction

Over the years, the expansion of IT and numerous technologies in medical imaging scanners has made room for their evolution into powerful new devices used for diagnostics and interventional radiology. The information generated by these products help with early detection and treatment of diseases, such as cardiology, neurology, oncology, orthopedics and surgery, resulting in significant improvements in patient care. A typical medical imaging and diagnostic equipment network is illustrated in the picture below.

This network can be broken down into subcategories:

- Modalities that capture and generate the images
- Picture Archiving and Communications Systems (PACS) that store generated images and make them available to physicians for diagnosis and treatments. It provides faster access to diagnostic information, reducing the need for film and film storage, in turn eliminating the problem of lost films and increases radiologist and physician satisfaction and productivity

- Radiology Information Systems (RIS) and Hospital Information Systems (HIS) that not only monitor and manage the workflow of radiology departments, but entire hospitals—from patient check-in to scheduling, billing and generating electronic medical records and management reporting
- Computed Radiography, which converts films to digital images
- Digital Radiography, which provides digital images
- Laser printers and other peripherals, which prints films when required

Modalities involve several scanning techniques to visualise the human body for diagnostic and treatment purposes, which include: Computed Tomography (CT), Magnetic Resonance Imaging (MRI), Angiography Therapies (AT), Positron Emission Tomography (PET) and Ultrasound (US). They are connected to PACS and RIS/HIS through Local Area Networks, Wireless Networks, or Wide Area Networks. PACS often have a dedicated Storage System to store all the diagnostic




Figure 1. Typical Medical Imaging & Diagnostic Equipment Network

imaging files into local servers or within a cloud platform. The main factors which are driving the fast deployment of these technologies are the needs of the hospitals to ensure compliance with regulations, control costs, and improve the quality of their services. The backbone of a digital hospital is a network made up of different modalities, PACS, RIS/HIS, Computed Radiography (CR)/Digital Radiography (DR), printers and peripherals. This complex network and its components must be compliant with several standards, like DICOM, Health Language Seven (HL7), the International Organization for Standardization (ISO), the International Electrotechnical Commission (IEC), the Body of European Regulators for Electronic Communications (BEREC), and other national codes such as BS7671:2001 (U.K.), CEI (Italy), and VDE (Germany), as well as the CEI IEC 60364 standard. Across most of Europe, Africa and the Middle East - with the notable exception of Germany and Austria - standard UPS systems must meet IEC standards for electrical equipment in healthcare settings, including critical life support systems. UPS systems are not, however, considered medical devices, which means they are subject to regulations limiting their proximity to patients. A UPS system cannot be connected to a patient and cannot be located within 1.5 meters of a patient. More relevantly, a UPS system must be installed upstream from the safety transformer. This transformer acts as a boundary between normal building power distribution and distribution for medical equipment. This physical infrastructure must be reliable, scalable, highly available and manageable.

It is composed of:

- **Power systems** such as UPS, Power Distribution Units (PDUs) and isolation transformers and generators to provide uninterrupted, clean power to critical loads
- Precision cooling systems that ensure proper environment conditions by adjusting temperature and humidity
- **Racks** that house the critical network equipment like servers, switches and routers, which host critical hospital applications
- Physical security and fire protection systems
- Cabling to interconnect equipment
- **Management systems** to monitor and manage the infrastructure, locally and remotely to ensure continuous operation
- **Technical services** to install, commission and maintain the systems

The hospital power system is a large and complex electrical network consisting of high voltage transformers, Automatic Transfer Switches (ATS), generators, isolation transformers, PDUs, etc. This system supplies power to electrical appliances, such as Heating Ventilation Air Conditioning (HVAC) systems, lighting, elevators, pumps, security and safety systems. The nature of these loads, which turn on and off randomly, creates an unstable power environment, creating potential sags and surges that sensitive imaging, diagnostic equipment and other IT devices must sustain. It is highly suggested for healthcare organisations to engage partners like Vertiv with a deep and global engineering expertise to assess the current status and suggest corrective actions.

Diagnostic and Treatment Applications

Modalities

Depending on the patient's disease, physicians can use different modalities for diagnosis and treatment (e.g. X-Ray or CT for orthopedics, ECG or MRI for cardiology, etc.). These modalities can be classified into two categories: **portable and stationary**. Portables can be further classified into hand-held (e.g. blood glucometer) or cart-mounted (e.g. ultrasound) while stationary devices can be further classified into desk-mounted (e.g. blood, urine analysis equipment) or floor-mounted (e.g. CT, MRI). The image below illustrates an MRI device.



Figure 2. MRI Device



In Vitro Diagnostic Laboratories

Clinical diagnostic instruments are used to detect health and disease state markers. Some of the key measurements that diagnostic analysers are used for are: metabolic markers (blood chemistry), liver function tests, heart disease and heart attack indicators, as well as viral and microbial infections such as HIV or H. pylori.

In Vitro Diagnostics (IVDs) are influencing over 60% of clinical decision-making, while accounting for about only 2% of total healthcare spend. With the development of Personalized Healthcare (PHC), the value of in vitro diagnostic laboratories plays a fundamental role in clinical decision-making, as patients can now benefit from customised treatments based on the analysis of genetic defects or biomarkers in their blood or tissue.

Targeted therapies and diagnostic tests help to improve medical decision-making with clear benefits for patients, while bringing economic advantages to healthcare organisations.

As electrical laboratory equipment may be sensitive to fluctuations in power supply, it is essential to protect all laboratory equipment by providing stable power through UPS. Certain types of equipment often need batteries as back up in case of power failure.

It is important to understand the sensitivity of this equipment to power outages and power spikes as these may affect its lifespan. For both the nature of service that these devices provide and related business model, which is often based on the "pay per result" approach, functionality must be ensured without disruption.

Field Technical Details

Modalities are generally used in an indoor environment. The cart-mounted and desk-mounted modalities generally use 230 VAC single-phase with less than 5 kVA of power, while floor-mounted devices typically require 400 VAC three-phase, ranging from 20 kVA to 300 kVA of power.

Floor mount, large imaging systems, such as Angiographers, can create a large power draw while operating under peak demands as high as 400 kW. With increased use of diagnostic equipment in most facilities, this can force system engineers to oversize the power system in order to compensate for these machines, resulting in higher initial costs and reduced efficiencies.

Similarly, the addition of new generation diagnostic scanners in a facility may increase the likelihood of a power overload condition, threatening the availability of the entire system or forcing an expensive upgrade. In response to this challenge, Vertiv has worked with medical equipment manufacturers to pioneer the use of UPS systems in peak shaving applications. With this technology, the critical load will be supported by both the utility source as well as the batteries, once it reaches the peak load. These peaks typically last less than 20 seconds and can occur sporadically throughout the day, depending upon the nature of the machine. Smart use of dedicated UPS systems upstream from diagnostic equipment allows the network to be sized according to "normal peaks" rather than to extreme conditions created by power-hungry units, ultimately reducing initial costs and improving operating efficiency. In addition, the UPS battery system can be used to protect against outages, as well as peak shaving, depending on how the battery is sized.



Figure 3. In Vitro Laboratory

Batteries require space adjacent the UPS. These are often placed in the technical room where all the systems needed to power, cool and control the imaging device are located. This room may be cooled with the building's comfort airconditioning system or a precision air conditioning system, which more tightly controls temperature and humidity in the environment. Considering that this room also hosts batteries to provide energy back up to critical equipment, an accurate temperature control is needed. According to the EN50272-2 standard*, in case of VRLA battery installation, a specific air quantity shall be supplied to the room to ensure proper ventilation. When a hospital's centralised chilled water pipeline is not available in the dedicated technical room, the deployment of a chiller to feed the MRI cooling system must be planned. The average power of this chiller should range from 40 kVA to 70 kVA depending on the imaging system power and whether the chiller is used for other purposes.

Hybrid Operating Theatres

A revolutionary alternative to conventional operating rooms is represented by the hybrid Operating Room (OR), which allows physicians to perform procedures using real-time image guidance and manage perioperative complications, all in a single environment. A hybrid operating room is an operating theatre equipped with a large fixed imaging system that supports high-quality interventional imaging as well as complex open procedures and minimally invasive surgeries.

Among the drivers that are pushing the development of such infrastructure, physicians desire to perform new combinations of endovascular, laparoscopic and/or open

*The EN 50272-2 standard will soon be replaced by standard EN IEC 62485-2.

procedures in the same OR using advanced angiographic image guidance. As new and emerging endovascular procedures are becoming more complex and of high risk, the patient benefits of reduced trauma/faster recovery associated with minimally invasive surgery and interventional procedures. However, there are also challenges facing an implementation of a hybrid OR such as high costs (about \$3-4 million on average), implementation time (1-2 years from planning to implementation), space requirements (80-130 sqm - close to double the size of a standard OR), staff training and team development.

Technical Recommendations

Since the hospital power grid is electrically "noisy and dirty", with many electrical surges and sags, it is good practice to provide UPS protection to all sensitive, expensive electronic systems.

UPS systems protect hardware, avoid unwarranted system crashes while tests are in progress, prevent loss of patient data files, and provide safe, reliable radiology examinations.

Electrical power systems used in the hospital should meet several standards, including:

- **IEC 60364-7-710:** Electrical installation requirements for medical locations
- **EN 62040-3:** Electromagnetic compatibility for UPS systems
- EN 50272-2*: Standard for batteries
- ISO 8528-1: Standard for gen-set

In Group 2, rooms (medical locations where applied parts are used in applications such as intracardiac procedures,



Figure 4. Hybrid Operating Room



operating theatres and vital treatment where failure of the supply can cause danger to life), the use of unearthed systems (IT*) is a requirement, as an interruption to the power supply is potentially life-threatening to the patients being treated. Treatments cannot be interrupted or repeated and test results need to be saved. The mandatory use of the IT system means the medical electrical devices/systems remain active on a first fault and there is no dangerous increase in the touch voltage. The distribution system should facilitate the automatic changeover from the main distribution system to the electrical safety power source which feeds the essential loads. This automatic changeover device requires a "safe separation" between systems:

- Luminaires of operating theatres, tables and other essential luminaires need a power changeover period of <0.5 secs. for a minimum autonomy of 3 hours;
- Safety lighting, medical electrical equipment in Group 2 medical locations, equipment of medical gas supply, and fire detection need a power changeover period of <15 secs.;
- Equipment essential for maintaining hospital services (i.e. cooling equipment, cooking equipment, sterilisation equipment) need a power changeover period of >15 secs.

Depending upon their usage, cart-mounted and deskmounted devices may require the UPS to comply with the international standard IEC60601-1 for patient vicinity applications. For large floor-mounted modalities, a UPS typically from 50- 300 kVA should be installed to protect the infrastructure. In cases where a large UPS cannot be installed, a smaller one (5-10 kVA), dedicated to the sensitive electronics and computer systems of non lifecritical modalities, like CT, MRI, and PET, should also be considered.

As discussed above in the "Field Technical Details" paragraph, sizing of the UPS for many devices, such as CTs and MRIs, can be challenging as they draw very high amounts of inrush current, thus great precautions should be taken while sizing their power systems (including UPS, generators, transformers, and switchgears). Their normal power consumption, as well as the inrush current ratings, are available from their modality manufacturers. It is important to allow enough margin for miscellaneous loads and future growth.

Adequate cooling and airflow should be provided for all modalities that have sensitive electronics dissipating heat. For most of the cart and desk-mounted modalities, building HVAC should be sufficient, however, for large floor-mounted modalities like CTs, MRIs, or PETs, supplemental cooling may be required. Precision cooling is preferred as it can provide temperature and humidity control in the CT/MRI room. All of the networked modalities and their physical infrastructure should be monitored and managed (i.e. environmental conditions of the radiology room, UPS battery life, runtime and capacity, and generator fuel) so that anomalies can be quickly detected, and corrective action can be proactively taken to avoid downtime.



*Insulated System IEC 60364-1 and IEC 61557-9 I: All live parts isolated from earth or with one point connected to earth through an impedance T: direct electrical connection of the exposed-conductive parts to earth

Picture Archiving and Communication System (PACS)

Field Technical Details

PACS (Picture Archiving and Communication System) is a healthcare technology for the short and long-term storage, retrieval, management, distribution and presentation of medical images.

A PACS allows a healthcare organisation to capture, store, view and share all types of images internally and externally. When deploying a PACS, a healthcare organisation must consider the environment in which it will be used and the electronic systems which it will be integrated with.

These systems are made up of a broad range of technologies that enable digital radiology and hospitals to perform teleradiology, telemedicine, and telesurgery. Diagnostic images will be available anytime, anywhere, making their distribution faster, easier, and more reliable. The core of the PACS is made up of storage and server clusters, housed in racks, computers or data centre environments. Typically, PACS need less than 10 kVA, single-phase AC power at 230 VAC.


PACS need to be available on demand for physicians and specialist surgeons, providing latest imaging data of the patient under treatment. It needs to be available at all times and there is little tolerance for downtime. Since the server clusters are contained within rack enclosures, handling their heat dissipation within the racks often becomes a challenge.

Technical Recommendations

PACS should be protected with an N+1 redundant UPS system which protects hardware and software from malfunctioning, and enables smooth shut downs, rebooting operations if necessary so as to prevent system crashes. For smaller, simpler systems, a basic UPS is sufficient. Considering that, often, additional outlets are needed to plug in all required devices, rack-based PDUs should be used for this scope. PDUs that can measure and display the absorption of current, which can help prevent accidental overloading and enable the PACS shut down, are recommended. For workstations running software applications, UPS protection with smooth shut down and reboot capabilities is recommended. PACS storage and servers should be housed in lockable rack enclosures.

Racks housing PACS storage and servers are generally very dense in terms of space and power consumption, and should thus be placed in a temperature-controlled environment.

Moreover, they should also have perforated doors for maximum airflow. When power absorption within the rack exceeds 4 kW, a row-based cooling unit can be used to provide supplemental cooling capacity. A good strategy involves the management of PACS servers, storage, and their entire physical infrastructure including UPS, PDUs, batteries and their critical environment (temperature and humidity). This will give early warnings of anomalies or impending disasters so that corrective actions can be taken and shutdowns prevented.





Radiology Information Systems (RIS) and Hospital Information Systems (HIS)

RIS and HIS are server-based systems running special software that make it possible to store, monitor, manage, and distribute patient medical information. They help patients in scheduling appointments, registration and billing, and support hospitals in generating, maintaining, and managing patients' electronic medical records, as well as generating workflows, worklists, management reporting, and a variety of other tasks. These RIS and HIS are becoming one large Hospital Information System and are integrated with PACS as well as various other modalities within the hospitals, providing complete automation. By converting them into "digital hospitals", they can significantly improve patient care, minimising human errors, saving lives, and reducing costs.

Field Technical Details

These systems are generally housed in a data centre environment, absorbing from 10 kW of single-phase 230 VAC power to 200 kW three-phase 400 VAC power.




Figure 8. Example of RIS and HIS Server-based Systems

Most data centres within hospitals have a UPS with battery back up, precision air conditioning units and a back up generator.

RIS / HIS are the most important systems within the data centre, requiring longer runtimes and higher redundancy and availability than most other equipment. Since the entire hospital depends on HIS for normal functionality, its availability requirements are generally 99.999%, which means an average unplanned downtime of 5 minutes per year or less. As for site preparation, attention should be paid to the floor load in terms of weight capacity, elevator capacity and door dimension to ensure that the physical infrastructure, like UPS, batteries, and air conditioning can be transported to their planned positions, as well as proper service maintenance, which should regularly and easily be done.

Technical Recommendations

The physical infrastructure supporting RIS / HIS should provide the highest levels of redundancy while minimising total cost of ownership. An N+1 redundant UPS with automatic and manual bypass is recommended, and sometimes this redundancy level is extended to the generator and precision air conditioning system to ensure the highest levels of availability. The entire infrastructure should be scalable to allow for future expansion and serviceability to reduce mean time to recover. Higher levels of redundancy, like dual-power feeds with dual generators and dual N+1 UPS with dual power paths all the way to the rack, should be considered for highly critical networks and data centres. PDUs should be able to measure and display current, which can help prevent accidental overloading and shutdown of the RIS / HIS. PDUs that allow remote outlet control via the web are desirable for rapid server rebooting. Isolation transformers should be used wherever required by local laws and precision air conditioning equipment should have the capability of allowing for expansion.

Conclusion

To ensure high availability and reliability to medical imaging and diagnostic equipment, including PACS, RIS, HIS, modalities, and their networks, special attention must be paid to their physical infrastructure. The biggest challenges are power continuity, cooling, physical space, management, and services. Providing UPS protection to all such devices protects hardware, prevents software crashes, and significantly increases their availability.

Cooling is a particular issue for larger floor-mounted modalities, high density storage and servers for PACS, as well as RIS / HIS and hospital wiring closets. In some cases, a building's HVAC system, along with proper ducting, ventilation, and airflow, may be enough. However, in many situations, additional cooling in the form of precision air conditioning is required. Companies like Vertiv have dedicated teams of system engineers who are experts in performing assessments of a data centre's physical infrastructure, and can provide detailed reports aimed at improving overall system reliability and availability while minimising total cost of ownership.





VertivCo.com | Vertiv Infrastructure Limited, George Curl Way, Southampton, SO18 2RY, VAT Number: GB188146827

© 2018 Vertiv Co. All rights reserved. Vertiv[™], the Vertiv logo and Liebert[®] GXT4 are trademarks or registered trademarks of Vertiv Co. All other names and logos referred to are trademarks or registered trademarks of their respective owners. While every precaution has been taken to ensure accuracy and completeness herein, Vertiv Co. Assumes no responsibility, and disclaims all liability, for damages resulting from use of this information or for any errors or omissions. Specifications are subject to change without notice.